



EU Quality Management Certificate



This is to certify that the company

SprintRay Inc.

2710 Media Center Drive, Suite #100A
Los Angeles, CA 90065
United States of America

SRN: US-MF-000036081

has established, implemented and maintains a Quality Management System in accordance with

Annex IX, Chapter I and III of the Regulation (EU) 2017/745 Conformity Assessment based on a Quality Management System and on Assessment of Technical Documentation

for the device categories and products listed in the Annex of this certificate.

The conformity of the Quality Management System has been verified in an audit and is subject to regular surveillance in accordance with Annex IX, Chapter 1, Section 3. Limitations to this certificate are listed in the Annex.

Devices listed in the Annex may bear the CE marking with the identification number of the Notified Body (0297).

For placing of devices of class III and devices class IIb implantable according to Article 52(4) subparagraph 2 listed in the Annex on the market, an additional certificate according to Annex IX, Chapter II is required.

Certificate registration no.	31622718 MDR2017Q
Certificate ID	1000302773
Effective date	2026-04-17
Expiry date	2030-04-29
Frankfurt am Main,	2026-04-17



DQS Medizinprodukte GmbH

Heinrich von Mettenheim
Managing Director



Accredited Body: DQS Medizinprodukte GmbH, August-Schanz-Str. 21, 60433 Frankfurt am Main
DQS Medizinprodukte GmbH is a Notified Body according to Regulation (EU) 2017/745 of the Council concerning medical devices with the Identification Number 0297.
The validity of the certification can only be verified by the QR-code.



Annex to EU Quality Management Certificate
SRN of Manufacturer: US-MF-000036081
Certificate ID: 1000302773

Authorised Representative of the company:

SprintRay Europe GmbH

Brunnenweg 11
64331 Weiterstadt
Germany

SRN: DE-AR-000033436

Device categories and variants covered by this certificate according to Article 52:

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: OnX Tough 2
Risk classification: IIa
Basic-UDI-DI: 0850039704SRI-0202164HG
Intended purpose: SprintRay OnX Tough 2 is a tooth shade ceramic-hybrid resin used for the fabrication of hybrid denture prosthetics, implant-supported denture prosthetics, monolithic full and partial removable dentures, and preformed denture teeth to be used in a denture.

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: Ceramic Crown
Risk classification: IIa
Basic-UDI-DI: 0850039704SRI-0202057HE
Intended purpose: SprintRay Ceramic Crown is a light-curable polymerizable resin intended to be used for the fabrication of; individual and fixed definitive full single crowns; definitive partial crowns in anterior and posterior area, individual and fixed single veneers; artificial teeth for dental prostheses, which are used for removable definitive full dentures; and individual and removable monolithic full and partial dentures in dental offices and laboratories. The material is an alternative to traditional restorative dental material.

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: APEX Base
Risk classification: IIa
Basic-UDI-DI: 0850039704SRI-0202153HB
Intended purpose: APEX Base resin is a light-curable polymerizable resin intended to be used for the fabrication and repair of full and partial removable dentures and baseplates. The material is an alternative to traditional denture base material.

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: APEX Teeth
Risk classification: IIa
Basic-UDI-DI: 0850039704SRI-0202157HK
Intended purpose: APEX Teeth is a light-curable polymerizable resin intended to be used to fabricate artificial teeth for full and partial removable dentures. This material is an alternative to traditional denture teeth material.



Annex to EU Quality Management Certificate
SRN of Manufacturer: US-MF-000036081
Certificate ID: 1000302773

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: Crown HT
Risk classification: IIa
Basic-UDI-DI: 0850070544SRI-0204030D3
Intended purpose: Crown HT is a light-curable polymerizable resin intended to be used for the fabrication of; individual and fixed definitive full single crowns; definitive partial crowns in anterior and posterior area, individual and fixed single veneers; artificial teeth for dental prostheses, which are used for removable definitive full dentures; and individual and removable monolithic full and partial dentures in dental offices and laboratories. The material is an alternative to traditional restorative dental material.

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: APEX Flex
Risk classification: IIa
Basic-UDI-DI: 0850070544SRI-0202187DP
Intended purpose: SprintRay APEX Flex is an alternative to traditional thermoplastic material used for the fabrication and repair of partial dentures. It is intended exclusively for professional dental work.

Device category: **MDN 1209 - Non-active non-implantable dental materials**
Product name: Digital Temp
Risk classification: IIa
Basic-UDI-DI: 0850039704SRI-0204020H5
Intended purpose: SprintRay Digital Temp is a light-curable resin indicated for the fabrication of individual and fixed temporary full single crowns, temporary partial crowns, and temporary bridges. The material is an alternative to traditional restorative dental material.

Examinations and tests performed:

31622718_A215301MED_01 dated 2024-10-19

31622718_A215326MED_02 OnX Tough 2 dated 2025-04-11

Further conditions for or limitations to the validity of the certificate:

n/a

Reference to previous certificates:

Revision	Date of Issue	Certificate-ID	Description of change
01	2025-04-30	100186243	Change of Basic UDI-DI
02	2025-08-07	1000258589	Addition of products APEX Base, APEX Teeth, Crown HT
03	2025-11-17	1000267003	Addition of products APEX Flex, Digital Temp